

### **SUPRA-PUBIC CYSTOSTOMY SET “METALLIC CANNULA” (With Polyurethane Drainage Catheter)**

#### **Device Description:**

**AMECATH SUPRA-PUBIC CYSTOSTOMY SET “METALLIC CANNULA”** is a sterile, single use device, for adult and pediatric use.

It is a splittable cystostomy metallic cannula supplied with Polyurethane drainage catheters (J Type or Malecot tip Catheters) or with two-way Foley Drainage catheter “total silicon ” with their accessories.

Splittable cystostomy metallic cannula: it is split needles for suprapubic bladder puncture. The optimized design provides minimal splitting forces and easy handling.

Range of size for Cystocath cannula is 10Fr. To 24Fr. Supplied with catheter of size between 8Fr. To 20Fr.

Range of size for metallic cannula is 3.05mm to 5.1mm supplied with catheter of size between 2.0mm to 4.66mm (8Fr. To 14Fr.)

#### **AMECATH SUPRA-PUBIC CYSTOSTOMY SET “METALLIC CANNULA” and its accessories:**

**AMECATH SUPRA-PUBIC CYSTOSTOMY SET “METALLIC CANNULA”** is available in different designs and Kit configurations to cover all customer needs.

**Target patient populations:** Adults and Paediatric

**Intended user:** Health Care Professionals

#### **Device construction:**

##### **List of Accessories:**

- Splittable METALLIC cannula.
- Polyurethane Catheter (J type or Malecot tip) with metal straightener.
- Connection tube (12Fr or 16Fr).
- Fixation wings.
- Stopcock 3 ways.
- Scalpel# 11.

#### **Intended Use:**

**AMECATH SUPRA-PUBIC CYSTOSTOMY SET “METALLIC CANNULA”** with Polyurethane drainage catheters, is used for:

- Bladder Drainage in case of failure of urethral catheterization or contraindicated urethral catheterization.
- Supra-pubic diversion after some selected urethral surgery.

#### **Contraindications:**

- Acute Urinary Retention due to clot retention caused by undiagnosed bladder mass to avoid malignant fistula.
- Previous Supra-pubic Surgery.

### Warnings and Precautions:

- For single product and patient use only. Do not use catheter if any sign of product damage is visible.
- Do not re-use, reprocess or re-sterilize. Reuse may lead to infection and Pyrogenicity. Reprocessing or Re-sterilization may damage the catheter and affect its integrity which when re-used may lead to possible deterioration in health and safety of patients.
- Caution should be taken while inserting the cannula / catheter into the bladder to avoid tissue damage or internal hemorrhage
- Catheter should not be left inside the body for more than 30 days.
- The methods of application are variable and could be modified by the HCP according to his own experience. HCP must be experienced or trained on the proper use of the device.
- The proper size selection for the catheter size and length is the responsibility of the physician.
- It is not recommended to use antimicrobial ointments on catheters as it may cause its degradation.
- Do not use absolute alcohol or acetone-based product on the catheter. 2% chlorhexidine or Iodine based solution is recommended as antiseptic solution
- The proper size selection for the device size and length is the responsibility of the physician considering the patient's anatomy.

### Complications:

- Sepsis
- Bleeding
- Pain
- Urine extravasation
- Encrustation

### How Supplied:

- Each **AMECATH SUPRA-PUBIC CYSTOSTOMY SET "METALLIC CANNULA"** is packed in a PETG hard blister covered with Tyvek.
- Window pack is applied when **AMECATH SSUPRA-PUBIC CYSTOSTOMY SET "METALLIC CANNULA"** is packed in a PETG hard blister with its PETG tray cover without tyvec lid.
- Each carton box includes 10 **AMECATH SUPRA-PUBIC CYSTOSTOMY SET "METALLIC CANNULA"**.

### **AMECATH SUPRA-PUBIC CYSTOSTOMY SET "METALLIC CANNULA" method of application** with polyurethane drainage catheter (J Type or Malecot tip Catheters)

1. Prepare the catheter before application by advancing the straightener inside the catheter to its end.
2. Choose the site of supra-pubic incision.
3. After applying the local anesthesia to the selected site, a small 2 cm. incision is done including the skin and subcutaneous tissue.
4. Introduce the metallic trocar cannula through the incision until it becomes inside the bladder, urine will come out through the trocar.

5. Advance the catheter with the straightener, through the trocar, urine will come through the catheter.
6. Remove the straightener.
7. Strip the metallic trocar cannula from the catheter.
8. Apply the fixation wing to the catheter and fix it to the skin.
9. Apply the extension tube to the catheter end and connect it to the urine bag.

### Catheter removal and exchange:

- Disconnect the urine bag, connection tube and fixation wings.
- Advance the straightener inside the catheter to the tip end to make it straight.
- Gently withdraw the catheter.
- Apply sterile dressing as needed.
- In long standing cases (> 1 month) the tract, left, is safe to introduce a new catheter through it.

**N.B for further information on luer connections, please refer to latest version of BS EN ISO 80369-7**

### CATHETER CARE /MAINTENANCE:

- Simple dressing to avoid contact to the skin is recommended
- Catheter should be dependent with frequent checking of patency
- Always ensure good hand hygiene is performed prior to any intervention and use protective equipment e.g., gloves
- Observe the cystostomy site for signs of infection and over granulation.

..

### Product Variants:

For variants of **AMECATH SUPRA-PUBIC CYSTOSTOMY SET “METALLIC CANNULA”**, Kindly refer to the catalogue, visit our website on: “[www.amecathgroup.com](http://www.amecathgroup.com)”, or contact your nearest **AMECATH** representative.”

Reference	Fr. Size	Contents
SUP-XX-J-K	XX	<ul style="list-style-type: none"> <li>● Splitable Cystostomy Metallic Cannula.</li> <li>● Catheter J Type (Pigtail) XXFr with Straightener.</li> <li>● Fixation Wings.</li> <li>● Connection Tube.</li> <li>● Stopcock 3 Ways.</li> <li>● Scalpel #11</li> </ul>
SUP-XX-M-K	XX	<ul style="list-style-type: none"> <li>● Splitable Cystostomy Metallic Cannula.</li> <li>● Catheter Malecot Type XXFr with Straightener.</li> <li>● Fixation Wings.</li> <li>● Connection Tube.</li> <li>● Stopcock 3 Ways.</li> <li>● Scalpel #11</li> </ul>

### Storage and Product Safe Disposal

- Store between 5°C to 30°C.
  - Do not expose to organic solvents, ionizing radiation or ultraviolet light.
  - Rotate inventory so that catheters are used prior to expiration date on the package label.
  - Used product should be disposed in sanitary container to prevent possible contamination and cross infection.
- ☐ N.B. please provide patients by the instructions of home care attached to this document.
- ☐ In case of any questions or queries, Kindly contact the local Authorised Representative or visit **AMECATH** website on : “ [www.amecathgroup.com](http://www.amecathgroup.com) ”
- ☐ In case of any Adverse event, Contact your local Health Authority immediately.
- ☐ Any serious incident that has occurred in relation to the device should be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established.



**AMECO MEDICAL INDUSTRIES**  
 Industrial Zone B4 - Plot T19 East.  
 10th of Ramadan City - Egypt  
 Tel: +20 5545 01321/2 - Fax: +20554501224  
[support@amecathgroup.com](mailto:support@amecathgroup.com) | [www.amecathgroup.com](http://www.amecathgroup.com)



**OBELIS S.A**  
 Bd. Général Wahis, 53  
 1030 Brussels - BELGIUM  
 Tel: +322.732.59.54 | Fax: +322.732.60.03  
[mail@obelis.net](mailto:mail@obelis.net) | [www.obelis.net](http://www.obelis.net)